



INSTRUMENT CLEANING AND PRE-DISINFECTION

Anios'Clean Excel D

FOR THE CLEANING AND DISINFECTION OF SURGICAL AND MEDICAL INSTRUMENTS

- Patented concept
- Original combination of active substances in synergy
- High level of detergence demonstrated
- Non-corrosive, compatible with a broad range of materials



Anios'Clean Excel D

For the cleaning and disinfection of surgical and medical instruments


PRODUCT CHARACTERISTICS

- Detergent and disinfectant designed to reduce the risk of surface biofilm formation on medical devices
- Broad antimicrobial spectrum to limit the risk of cross-contamination and ensure the safety of personnel
- Antimicrobial efficacy demonstrated, including on MRB and HRB
- Formulation active in fresh and hard water
- Formulation non-corrosive and compatible with a broad range of materials
- Improvement of the ecotoxicological profile:
GHG* emissions reduced by 47% relative to a similar product
- Good rinseability
- Can be used in an ultrasound cabinet
- Clear blue solution
- Pure product pH: min. 6.5 - max. 7.5
- pH at 0.5% in mains water: 7.4
- Lemon fragrance

INSTRUCTIONS FOR USE




Dilution at 0.5%: Pour a 25 ml dose per 5 litres of cold or tepid water. Renew the soaking bath once a day at least.



Totally immerse medical device. Soaking time: 15 minutes. Brush if visible soiling is present. For endoscopic equipment: brush channels.



Thoroughly rinse treated medical devices with tap water (good microbiological quality) to remove all traces of product and organic soiling. For endoscopic equipment: rinse internal and external parts.



Dry with a clean towel. Proceed to the next step (see protocol established by the service).

MICROBIOLOGICAL PROPERTIES

CLAIM/CONDITIONS	CONCENTRATION %	CONTACT TIME
Bactericidal According to EN 13727 and EN 14561 According to EN 13727 against MdRB	0,5 0,5	5 min 15 min
Mycobactericidal According to EN 14348 and EN 14563 (M. terrae)	2,0	60 min
Yeasticidal According to EN 13624 and EN 14562	0,5	15 min
Virucidal PRV (surrogate of HBV), BVDV (surrogate of HCV), Herpesvirus, Vaccinia virus, HIV-1	0,5	5 min

PACKAGING AND ORDERING INFORMATION

PRODUCT	PACKAGING UNIT	ORDER CODE
Anios'Clean Excel D	200 x 25ml doses	XXXXXX
Anios'Clean Excel D	12 x 1L bottle	XXXXXX
Anios'Clean Excel D	4 x 5L can with 25ml pump	XXXXXX



ANIOS
Pavé du Moulin
59260 Lille-Hellemmes
0459 France

Distributed by:
ECOLAB DEUTSCHLAND GmbH
Ecolab-Allee 1
40789 Monheim am Rhein
Germany

APPLICATION AREA

Cleaning and disinfection of medical devices: surgical and medical instruments, heat-sensitive instruments and endoscopy equipment, before sterilisation.

INSTRUCTIONS FOR USE

1. Dilution at 0.5%: Pour a 25 ml dose per 5 litres of cold or tepid water. Renew the soaking bath once a day at least.
2. Totally immerse medical device. Soaking time: 15 minutes. Brush if visible soiling is present. For endoscopic equipment: brush channels.
3. Thoroughly rinse treated medical devices with tap water (good microbiological quality) to remove all traces of product and organic soiling. For endoscopic equipment: rinse internal and external parts.
4. Dry with a clean towel. Proceed to the next step (see protocol established by the service).

INGREDIENTS

Didecyldimethylammonium chloride, chlorhexidine digluconate, non-ionic surfactants, perfume, colouring, excipients.

STORAGE

Dangerous - respect the precautions for use (drawn up according to the European rules in force regarding the classification and labelling of chemical products). Storage: between +5°C and +35°C.

CERTIFICATES

Medical device. Full antimicrobial activity for Anios Clean Excel D is included in the scientific dossier, available on request.

MEDICAL DEVICE USE SAFELY

Please read label and product information.



FOR PROFESSIONAL USE ONLY.

ECOLAB®

www.ecolab.com